

Certificate

Traceability of Manufacturers to the IFCC Reference Measurement Procedure for HbA1c

This certifies that *Polymer Technology Systems Inc.* using *Tosoh G7*, participates in the Monitoring Programme to demonstrate traceability. In the Monitoring Programme of 2015 the following performance was seen:

Deviation from IFCC-target	at 30 mmol HbA1c/mol Hb :	0.3
	at 60 mmol HbA1c/mol Hb :	0.8
	at 90 mmol HbA1c/mol Hb :	1.2
Reproducibility, coefficient of variation		1.30%
Linearity, correlation coefficient		0.9985

Date of issue: 4 December 2015

Certification expires: 31 December 2016



Dr. C.W. Weykamp
IFCC HbA1cNetwork Coordinator